



I agree to:

1. Use this machine responsibly.
2. Take full responsibility for decisions, care and use of the Sentient Element.
3. Read user manual and watch any relevant "how to" tutorial videos made by Sentient Light.
4. Read the terms and conditions.
5. Read the money back guarantee policy.
6. Never store coils in the Sentient Element case.
7. Have any kind of liquids on or near the machine.
8. Keep the coils 1-2ft away from any electronics, such as laptop or smartphones, while using the machine.
9. Save the shipping box for 6 months.
10. Start with low amplitude/volume at the beginning of all PEMF sessions.
11. Decrease power if the machine trips off more than once. The machine will trip off if the amplifier overheats. This is an indicator that you are running it too high. Please turn it down.
12. Not overheat the coil. Do not run it on full power for more than 5-10 mins. Make sure to have a 30 min cool down time, as it can burn out the amplifier and over heat the coils.
13. Never sleep with coil or leave unattended for a long period of time. It can burn or melt certain materials. We recommend leaving the coils in cover.
14. Do not put coils or heavy items inside the Sentient Element case.
15. Always have adult supervision when using it on or around younger children.
16. Never use it on yourself (or anybody else) if you have a pacemaker, insulin pump or any other kind of electrical implant. The danger is that the electromagnetic pulses used in PEMF therapy could interfere with the device and cause failure.
17. Used with caution or with professional guidance if you have implanted metals, such as joint replacements, dental implants, mechanical heart valves, metal stents, or metal staples in blood vessels. This is not due to risk of injury or harm, but because extremely high intensity PEMFs may stimulate the nerves in the area of the metal, causing sharp pain. High intensity PEMFs may also add to the shear stress of metal clips placed in or near blood vessels.
18. Do not use high intensity PEMFs over breast implants. High-frequency PEMFs beyond 100 Hz is probably also not desirable for treatment durations longer than an hour at a time, given the risk of agitating the plastic or silicone in breast implants, resulting in possible thinning and risk of leakage.
19. Contact Sentient Light if there are any technical issues with the Sentient Element device or coils.
20. Never unscrew and/or open the top panel of the Sentient Element device. Warranty becomes void if you do.
21. Never make modifications with any part or piece of the Sentient Element device and/or coils.
22. Follow instructions and warnings on the label located inside the case lid.

Read Disclaimer: This product is not meant to treat, prevent any disease, or replace proper medical attention. If you are pregnant, taking medication, have a medical condition consult your physician before using this product. This product has been known to trigger herxheimer reaction (body's elimination process to rid itself of toxins and waste) and PEMF should be used with caution. We are not responsible for any repercussions for use outside of the specifications covered in this manual.